


Untitled

Instrument


Instrument type	2255
Instrument serial number	100043
Instrument software type	FW-2255-000
Instrument software version	1.2.0.1151
Transducer type	4966
Transducer serial number	3352521
Sound field	Free-field
Windscreen	UA-1650
Calibration date	31/10/2023

Measurement



Type	ID	Name	Start time	End time	Elapsed time	LAeq [dB]	LCeq [dB]	LZeq [dB]	LZeq(31.5 Hz - 500 Hz) [dB]	LAFmax [dB]	LASmax [dB]	LAFmin [dB]	LCpeak [dB]	LZpeak [dB]	LAE [dB]	LAF 1.0 [dB]	LAF10.0 [dB]	LAF50.0 [dB]	LAF90.0 [dB]
Total	Sum				00:15:00	63.31	67.30	72.07	66.06	73.32	67.19	53.26	95.31	96.11	92.85	66.98	65.41	63.00	60.05
Total-Exclude	Sum				00:15:00	63.31	67.30	72.07	66.06	73.32	67.19	53.26	95.31	96.11	92.85	66.98	65.41	63.00	60.05
Cursor	01	2255-100043 01	13/06/2025 7:52:20 AM	13/06/2025 7:52:30 AM	10	65.05	68.83	74.22	67.59	68.15	66.48	62.58	82.60	88.35	75.05	67.70	66.30	65.06	63.40

Spectrum



Frequency [Hz]	Total 2255-100043 01 LAeq [dB]
16	6.85
31.5	22.25
63	36.67
125	43.02
250	47.05
500	53.96
1k	61.28

2k	56.28
4k	44.97
8k	29.93
16k	14.07